

Service Report 1023762_01

| | | | |
|------------------------|---|--------------------------|------------------------------|
| Customer | 18113159 - Statens Serum Institut | Service Call Type | Warranty |
| Address | Tyge Soendergaards Vej 951 8200 Århus N Denmark | Event Code | Preventive Maintenance Visit |
| General Contact | Anja Louise Wittrup O'Brien | | |
| Phone | +45 3268 5380 | | |
| Email | alwo@ssi.dk | | |



| | |
|-------------------|---|
| Instrument | STAR M Autoload, 8x 1000ul ch, CO-RE 96 MPH TADM, UV... |
| Serial No. | 745G |
| Request | Preventive Maintenance Visit |

| Working Time | Qty |
|------------------------------|------|
| Preventive Maintenance Visit | 5.00 |

| Travel Zone |
|-------------|
| Zone 1 |

| Spare Parts | Item No. | Batch No. | Qty |
|-----------------------|----------|-----------|------|
| CO-RE O-RING SET OF 8 | 183060 | | 1.00 |
| SET OF 8 STOP DISK PX | 183061 | | 1.00 |

| Work Description |
|--|
| 1023762_01 Niels Christian Brønderup : All Rails and spindles cleaned and relubricated All stop disks and O rings exchanged All macros completed succesfully Verification completed succesfully |

| | | | |
|-------------------------|---|---------------------------------|---|
| Customer On-site | Line nielsen | Field Service Technician | Niels Christian Brønderup |
| Date | 14.11.2022 | Date | 14.11.2022 |
| Signature |  | Signature |  |

Hamilton Nordic AB

Kistagången 20B
164 40 Kista
Sweden