

## Temperature Verification Report

<b>Instrument Name:</b>	ML_STAR	<b>Date:</b> 2022-11-14
<b>Instrument Serial No.:</b>	745G	<b>Time:</b> 13:45
<b>Instrument User Software Version:</b>	4.5.0.5217 FVK2: 1020/03	
<b>Location:</b>	SSI Vest	<b>Temperature [°C]:</b> 20.7
<b>Operator:</b>	Niels Brønderup	
<b>Reason for Verification:</b>	PM2	
<b>Report File:</b>	Report_Temperature_Ver_745G_202211141345.pdf	

**Test Equipment:**

<b>Temp. Measuring Device Serial No.:</b>	7032317	<b>Valid until:</b> 2023-02-24
---	---------	--------------------------------

**Process Status:** *passed*

**Operator:** NCB      **Date:** 14-11-2022      **Signature:** *Niels Brønderup*

Measurements:	Site 1	Site 2	Site 3	Site 4	Site 5
<b>Device Name / ID:</b>	Hamilton Heater Shaker	Hamilton Heater Shaker	Hamilton Heater Shaker	Hamilton Heater Shaker	Hamilton Heater Shaker
<b>Device Serial Number:</b>	219I	193I	216I	218I	217I
<b>Position (x / y [mm]):</b>	250 / 313	251 / 193	252 / 73	408 / 313	408 / 193
<b>Set temperature [°C]:</b>	60.0	60.0	60.0	60.0	60.0
<b>Temperature Offset [°C]:</b>	-0.2	-0.2	-0.2	-0.2	-0.2
<b>Target temperature [°C]:</b>	59.8	59.8	59.8	59.8	59.8
<b>Deviation (spec.) [°C]:</b>	<= +/- 2.0	<= +/- 2.0	<= +/- 2.0	<= +/- 2.0	<= +/- 2.0
<b>Measured temperature [°C]:</b>	60.6	59.8	59.9	60.0	60.3
<b>Test Result:</b>	passed	passed	passed	passed	passed
<b>Date of verification:</b>	2022-11-14	2022-11-14	2022-11-14	2022-11-14	2022-11-14

