

Position Verification Report for 1000 µl Pipetting Channels

Instrument Name:	ML_STAR	Date: 2022-11-14
Instrument Serial No.:	745G	Time: 12:36
Instrument User Software Version:	4.5.0.5217 FVK2: 1020/02	
Location:	SSI Vest	
Operator:	Niels Brønderup	
Reason for Verification:	PM2	
Report File:	Report_XYZ_Ver_1000ulSingleChannels_745G_202211141236.pdf	

Process Status: *passed*

Channel No.	Specification (mm)	Reproducibility in x-Direction (mm)	Reproducibility in y-Direction (mm)	Reproducibility in z-Direction (mm)	Status
1	<= +/- 0.3	0.0	0.0	0.0	passed
2	<= +/- 0.3	0.0	0.0	0.0	passed
3	<= +/- 0.3	0.0	0.0	0.0	passed
4	<= +/- 0.3	0.0	0.0	0.0	passed
5	<= +/- 0.3	0.0	0.0	0.0	passed
6	<= +/- 0.3	0.0	0.0	0.0	passed
7	<= +/- 0.3	0.0	0.0	0.0	passed
8	<= +/- 0.3	0.0	0.0	0.0	passed

Operator: NCB

Date: 14-11-2022

Signature: *Niels Brønderup*