

# Service Report

Case no.: 20829-T1Q2V3

Date: 05-12-2022

Statens Serum Institut

Att: Allan Povlsen  
 Artillerivej 5  
 2300 København S  
 DK

Phone: 32 68 32 68	Requisition/order no.:	Email: ALPO@ssi.dk
--------------------	------------------------	--------------------

Instrument: Biomek i7 MC	Location: TCDK vest
Contract:	Serial no.: B8758720L006
Date of Installation: 07-12-2020	
Reported error: PM 2023	

Date:	Start:	End:	Description	Hours
16-01-2023	12:00	14:00	PM 2023  Service eftersyn udført. Instrument efferset og smurt. 4-post testet - OK. Funktionstest udført - OK.  PM serviceeftersyn udført efter Beckman Coulter's Preventive Maintenance check.	2
Hours in Total				2

Item number	Name	Quantity	Wearing

Completed / delivered by: Service Engineer  Date: 16-01-2023

Signature

