

Position Verification Report

for 1000 µl Pipetting Channels

Instrument Name:	ML_STAR	Date: 2022-12-07
Instrument Serial No.:	403F	Time: 12:01
Instrument User Software Version:	4.5.0.5217 FVK2: 1020/02	
Location:	SSI TCDK	
Operator:	Fredrik Svensson	
Reason for Verification:	PM2	
Report File:	Report_XYZ_Ver_1000ulSingleChannels_403F_202212071201.pdf	

Process Status: passed

Channel No.	Specification (mm)	Reproducibility in x-Direction (mm)	Reproducibility in y-Direction (mm)	Reproducibility in z-Direction (mm)	Status
1	<= +/- 0.3	0.0	0.0	-0.1	passed
2	<= +/- 0.3	0.0	0.0	0.0	passed
3	<= +/- 0.3	0.0	0.0	0.0	passed
4	<= +/- 0.3	0.0	0.0	-0.1	passed
5	<= +/- 0.3	-0.1	0.0	0.0	passed
6	<= +/- 0.3	0.0	0.0	0.0	passed
7	<= +/- 0.3	0.0	0.0	0.0	passed
8	<= +/- 0.3	0.0	0.0	0.0	passed

Operator: Fredrik Svensson **Date:** 2022-12-07

Signature: *Frsu*