
Instrument Name:	ML_STAR	Date: 2022-12-07
Instrument Serial No.:	403F	Time: 12:33
Instrument User Software Version:	4.5.0.5217 FVK2: 1020/02	
Location:	SSI TCDK	
Operator:	Fredrik Svensson	
Reason for Verification:	PM2	
Report File:	Report_XYZ_Ver_96erPipHead_403F_202212071233.pdf	

Process Status: ***passed***

Specification (mm)	Reproducibility in x-Direction (mm)	Reproducibility in y-Direction (mm)	Reproducibility in z-Direction (mm)	Status
<= +/- 0.3	0.0	0.0	0.0	passed

Operator: Fredrik Svensson

Date: 2022-12-07

Signature: Frsv