

Service Report

Case no.: 20826-P5B4K2

Date: 05-12-2022

Statens Serum Institut

Att: Allan Povlsen
 Artillerivej 5
 2300 København S
 DK

Phone: 32 68 32 68	Requisition/order no.:	Email: ALPO@ssi.dk
--------------------	------------------------	--------------------

Instrument: Biomek i7 MC	Location: TCDK vest	
Contract:	Serial no.: B8758720M004	Date of Installation: 05-01-2021
Reported error: PM 2023		

Date:	Start:	End:	Description	Hours
20-02-2023	11:00	11:00	PM Biomek 2023 Service eftersyn udført. Instrument efferset og smurt. 4-post testet - OK. Funktionstest udført - OK. PM serviceeftersyn udført efter Beckman Coulter's Preventive Maintenance check.	0
Hours in Total				0

Item number	Name	Quantity	Wearing

Completed / delivered by: Service Engineer  Date: 22-02-2023

Signature

